



Republic of the Philippines
DEPARTMENT OF AGRICULTURE
 Regional Field Office No. 1
 Aguila Road, Sevilla, City of San Fernando, La Union
 Telephone No. (072) 242 1045/1046



SUPPLEMENTAL/BID BULLETIN NO. 01
Issued on 06 November 2024

Name of Project: NEGOTIATED PROCUREMENT OF SUPPLY, DELIVERY AND TESTING OF TWO (2) UNITS VETERINARY ULTRASOUND EQUIPMENT

IB Identification No.: DA-RFO 1-2024-GOODS-069-N

Reference No.: 11354243

This Supplemental/Bid Bulletin No. 01 is being issued to amend the **Particulars/Descriptions** of the **Technical Specifications** and **Delivery Period** as clarification to issues raised in the Preliminary Conference for the above-stated project conducted last 04 November 2024, to wit:

- I. Amendments to the **Particulars/Descriptions** of the **Technical Specifications** are hereby highlighted and underscored below:

ORIGINAL	AMENDED
Veterinary Ultrasound Equipment Scanning Method: Convex, Linear electronic scan Monitor: at least 8.4 inches Image Store: with Flash Memory Interface: USB port x 1 Power Source: Rechargeable Battery, 100V-240V AC 50/60Hz Probes: Linear Rectal Probe, Microconvex probe, Convex probe Accessories: Shoulder beltclip, AC adapter & Power supply cord, Charger base, Battery x 2, Gel (2 gallon), Probe Grip for endorectal probe, Manual of Operation, Protection cover and carrying case Result: real time, 2-D Bundle: ➤ xxx Additional requirement: During bid opening: ➤ Supplier/Bidder must have a valid FDA-issued License to Operate	Veterinary Ultrasound Equipment Scanning Method: Convex, Linear electronic scan Monitor: at least 8.4 inches Image Store: with Flash Memory Interface: USB port x 1 Power Source: Rechargeable Battery, 100V-240V AC 50/60Hz Probes: Linear Rectal Probe, Microconvex probe, Convex probe, <u>Cable Ranges 1.5 meters - 3 meters</u> Accessories: Shoulder beltclip, AC adapter & Power supply cord, Charger base, Battery x 2, Gel (2 gallon), Probe Grip for endorectal probe, Manual of Operation, Protection cover and carrying case Result: real time, <u>4-D</u> Bundle: ➤ xxx Additional requirement: During bid opening: ➤ Supplier/Bidder must have a valid FDA-issued License to Operate



Republic of the Philippines
DEPARTMENT OF AGRICULTURE
 Regional Field Office No. 1
 Aguila Road, Sevilla, City of San Fernando, La Union
 Telephone No. (072) 242 1045/1046



<p>Veterinary/Medical/Laboratory Devices importer/distributor</p> <ul style="list-style-type: none"> ➤ Attach brochure for the offered brand <p>During post-qualification:</p> <ul style="list-style-type: none"> ➤ 100% availability of the stocks during post-qualification <p>During delivery:</p> <ul style="list-style-type: none"> ➤ xxx 	<p>Veterinary/Medical/Laboratory Devices importer/distributor</p> <ul style="list-style-type: none"> ➤ Attach brochure for the offered brand ➤ <u>Certificate of Availability of spare parts beyond warranty</u> <p>During post-qualification:</p> <ul style="list-style-type: none"> ➤ <u>Duly notarized letter of Principal Supplier certifying the availability of stocks during the post-qualification</u> ➤ <u>Certified true copy of the Purchase Order of the stocks</u> <p>During delivery:</p> <ul style="list-style-type: none"> ➤ xxx
---	--

II. Amendment to the **Delivery Period** is hereby highlighted and underscored below:

ORIGINAL	AMENDED
at least forty-five (45) working days from the receipt of Notice to Proceed (NTP)	at least <u>twenty (20) working days</u> from the receipt of Notice to Proceed (NTP)

These amendments shall form part of the Bidding Documents and shall be made an integral part thereof.

For the prospective bidders' guidance, reference and information of all concerned.


GILBERT D. RABARA, DVM
 BAC Vice-Chairperson *GR*